
From: Clune, Bridget [JNJIE]
To: Flett, Magnus [DPYGB]; Kalra, Vishnu [DPYGB]
Sent: 6/10/2005 10:00:30 AM
Subject: FW: ASR revised today

Hi Guys,

I am quite concerned about our failure rate, I have had report from Emad Middle East, and Jose in Spain. All of these reports have been technique related, and the failure was in the seating of the cup. Have we had a history of failures over the last year or so or is this phenomenon related to the wider Roll-out?

Can we get together to discuss end next week?

Regards,

Bridget.

-----Original Message-----

From: Babidge, Jenny [DPYAU]
Sent: 09 June 2005 19:24
To: Flett, Magnus [DPYGB]; Clune, Bridget [JNJIE]; Kalra, Vishnu [DPYGB]
Cc: Collins, Troy [DPYAU]
Subject: FW: ASR revised today

Magnus, Vishnu, Bridget

We have done a few revisions now - am a little concerned that our number is growing. How is the rate globally?

I have not had official response for any retrieval we have sent to Hamburg. Only verbal on the one of Dr Zicat and Rogers

Regards

Jen

-----Original Message-----

From: Debbie Wilks [mailto:dcwilks@ozemail.com.au]
Sent: Thursday, 9 June 2005 9:55 PM
To: Babidge, Jenny [DPYAU]
Cc: Troy Collins; Mark Gibson; Loughlin, Marie [DPYAU]
Subject: ASR revised today

Hi Jen,

Just letting you know an ASR was revised today by Dr Zicat at the Mater Hospital. The original procedure was performed on the 21st October 2004 - implants used a 54 cup with a 47 head implanted with Endurance cement. The patient was revised due to increasing pain and presence of a gait. I spoke with Dr Zicat after the procedure and both implants were not loose but there was evidence of synovitis. The ASR cup was left insitu and the stem implanted was S-ROM. He has sent the sample for analysis to Dr Walsh at the POW lab as he feels he does not receive adequate feedback from the retrieved implants he has sent back to Depuy.

He will keep us informed of results he receives from Dr Walsh.

Regards,

Debbie.